

INTERNATIONAL STANDARD

**Reference conditions and procedures for testing industrial and process
measurement transmitters -
Part 2: Specific procedures for pressure transmitters**

CONTENTS

FOREWORD	2
INTRODUCTION	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 General description of the device and overview	9
5 Reference test conditions	9
6 Test procedures	10
6.1 General	10
6.2 Tests at standard and operating reference test conditions	10
6.2.1 General	10
6.2.2 Accuracy test suitable for routine and acceptance tests	11
6.2.3 Influence of the overpressure	12
6.2.4 Influence of static pressure	14
6.2.5 Long-term drift	16
6.2.6 Leakage test	17
6.2.7 Additional tests for diaphragm/remote seals – Influence of process temperature (long term)	18
6.2.8 Additional tests – Burst pressure	18
7 Test report and technical documentation	18
7.1 General	18
7.2 Total probable error	18
Annex A (informative) Relationship between the SI unit and other pressure related units	20
Annex B (informative) Pressure process measurement transmitter (PMT)	21
B.1 General description of a pressure PMT	21
B.2 Typical PMTs	21
Annex C (informative) Main characteristics for pressure transmitters	23
Bibliography	26
Figure 1 – Measuring range and associated properties of a pressure PMT	7
Figure 2 – Schematic example of a test set-up for pressure PMT	10
Figure 3 – Example of measured error plot	12
Figure 4 – Procedure for the determination of the unilateral overpressure error	13
Figure 5 – Schematic example of test set-up for determining the effect of the static pressure	14
Figure 6 – Procedure for the determination of the zero point error with static pressure	15
Figure 7 – Procedure for the determination of the span error for static pressure	16
Figure B.1 – Schematic example of intelligent PMT model	22
Table 1 – Example of measured errors with 20 % steps	11
Table A.1 – Relationship between the SI unit and other pressure related units	20
Table C.1 – Pressure-based transmitters	23

INTERNATIONAL ELECTROTECHNICAL COMMISSION

Reference conditions and procedures for testing industrial and process measurement transmitters - Part 2: Specific procedures for pressure transmitters

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62828-2 has been prepared by subcommittee 65B: Measurement and control devices, of IEC technical committee 65: Industrial-process measurement, control and automation. It is an International Standard.

This second edition cancels and replaces the first edition published in 2017. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) revision according to the latest IEC rules;
- b) comparison with the general part of IEC 62828-1;
- c) additions and more precise definitions in the "Terms and definitions" clause;
- d) correction of the calculation formulas for the measurement error;

- e) more precise formulation of long-term drift;
- f) extension of additional tests;
- g) transfer of the annex "Example of signal current range for a 4 to 20 mA PMT" in IEC 62828-1;
- h) new Annex C "Main characteristics for pressure transmitters".

The text of this International Standard is based on the following documents:

Draft	Report on voting
65B/1308/FDIS	65B/1319/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

This International Standard is to be used in conjunction with IEC 62828-1:2026.

A list of all parts in the IEC 62828 series, published under the general title *Reference conditions and procedures for testing industrial and process measurement transmitters*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

INTRODUCTION

Most of the current IEC standards on industrial and process measurement transmitters are rather old and were developed having in mind devices based on analogue technologies. Today's digital industrial and process measurement transmitters are quite different from those analogue transmitters: they include more functions and newer interfaces, both towards the computing section (mostly digital electronic) and towards the measuring section (mostly mechanical). Even if some standards dealing with digital process measurement transmitters already exist, they are not sufficient, since some aspects of the performance are not covered by appropriate test methods.

In addition, existing IEC test standards for industrial and process measurement transmitters are spread over many documents, so that for manufacturers and users it is difficult, impractical and time-consuming to identify and select all the standards to be applied to a device measuring a specific process quantity (pressure, temperature, flow, level, etc.).

To help manufacturers and users, it was decided to review, complete and reorganize the relevant IEC standards and to create a more suitable, effective and comprehensive standard series that provides in a systematic way all the necessary specifications and tests required for different industrial and process measurement transmitters.

To solve the issues mentioned above and to provide an added value for the stakeholders, the new standard series on industrial and process measurement transmitters covers the following main aspects:

- applicable normative references;
- specific terms and definitions;
- typical configurations and architectures for the various types of industrial and process measurement transmitters;
- hardware and software aspects;
- interfaces (to the process, to the operator, to the other measurement and control devices);
- physical, mechanical and electrical requirements and relevant tests; clear definition of the test categories: type tests, acceptance tests and routine tests;
- performance (its specification, tests and verification);
- environmental protection, hazardous areas application, functional safety, etc.;
- structure of the technical documentation.

To cover in a systematic way all the topics to be addressed, the IEC 62828 series is organized in several parts. At the moment of the publication of this document, the IEC 62828 series consist of the following parts:

- *IEC 62828-1: General procedures for all types of transmitters;*
- *IEC 62828-2: Specific procedures for pressure transmitters;*
- *IEC 62828-3: Specific procedures for temperature transmitters;*
- *IEC 62828-4: Specific procedures for level transmitters;*
- *IEC 62828-5: Specific procedures for flow transmitters.*

1 Scope

This part of IEC 62828 establishes specific procedures for testing pressure process measurement transmitters (PMT) used in measuring and control systems for industrial processes and for machinery.

A pressure PMT can feature a remote seal to bring the process variable to the sensing element in the PMT. When the remote seal cannot be separated from the PMT, the complete device is tested.

For general test procedures, reference is made to IEC 62828-1, which is applicable to all types of process measurement transmitters.

NOTE In industrial and process applications, to indicate the process measurement transmitters, it is common also to use the terms "industrial transmitters", or "process transmitters".

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 62828-1:2026, *Reference conditions and procedures for testing industrial and process measurement transmitters - Part 1: General procedures for all types of transmitters*

IEC 61518, *Mating dimensions between differential pressure (type) measuring instruments and flanged-on shut-off devices up to 413 bar (41,3 MPa)*

Bibliography

IEC 61010-1:2010, *Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements*

IEC 61010-1:2010/AMD1:2016

IEC 61987-13, *Industrial-process measurement and control - Data structures and elements in process equipment catalogues - Part 13: Lists of properties (LOP) for pressure measuring equipment for electronic data exchange*

IEC 62828 (all parts), *Reference conditions and procedures for testing industrial and process measurement transmitters*

IEC Common Data Dictionary (CDD): available at <https://cdd.iec.ch/>

EN 12266-1, *Industrial valves - Testing of metallic valves - Part 1: Pressure tests, test procedures and acceptance criteria - Mandatory requirements*
