

INTERNATIONAL STANDARD

CONSOLIDATED VERSION

**Electroacoustics - Audiometric equipment -
Part 6: Instruments for the measurement of otoacoustic emissions**

CONTENTS

FOREWORD	3
INTRODUCTION	5
INTRODUCTION to Amendment 1	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	7
4 Requirements for specific instruments	8
5 General specifications	8
5.1 Acoustic stimulus system	8
5.1.1 General requirements	8
5.1.2 Stimulus types	9
5.1.3 Stimulus frequency range	9
5.1.4 Stimulus level	9
5.1.5 Intermodulation distortion	10
5.2 Test quality assuring system	10
5.2.1 Stability of acoustic response in the external auditory meatus	10
5.2.2 Test quality assurance	11
5.2.3 Individual stimulus recordings	11
5.3 Measuring system	11
5.3.1 Units of measurement	11
5.3.2 Measurement range	11
5.3.3 Accuracy of measurement	11
5.3.4 Frequency range	11
5.3.5 Noise reduction	11
5.3.6 Response detection	11
5.3.7 Response quality estimates	11
5.3.8 Normative values	12
5.4 Presentation of results	12
6 Demonstration of conformity with specifications	12
6.1 General	12
6.2 Probe signal	12
6.2.1 Probe signal frequency spectrum	12
6.2.2 Probe signal level and harmonic distortion	12
6.2.3 Probe measurement accuracy	13
6.3 Complete system	13
6.4 Maximum permitted expanded uncertainty of measurements $U_{\max}$	13
7 General requirements	14
7.1 Marking	14
7.2 Instruction manual	14
7.3 Safety requirements	14
7.4 Immunity to power and radiofrequency fields	14
7.5 Warm-up time	14
7.6 Voltage supply variation and environmental conditions	15
7.6.1 Mains operation	15
7.6.2 Battery operation	15
7.6.3 Environmental conditions	15

8	Additional characteristics to be specified by the manufacturer	15
9	Periodic calibration	15
	Bibliography.....	16
	Table 1 – Mandatory functions for otoacoustic emission instruments	8
	Table 2 – Documentation of test conditions, parameters and results	12
	Table 3 – Values of $U_{\max}$ for conformance and periodic calibration measurements	14

INTERNATIONAL ELECTROTECHNICAL COMMISSION

Electroacoustics - Audiometric equipment - Part 6: Instruments for the measurement of otoacoustic emissions

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 60645-6 edition 2.1 contains the second edition (2022-03) [documents 29/1109/FDIS and 29/1114/RVD] and its amendment 1 (2026-08) [documents 29/1240/FDIS and 29/1242/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

IEC 60645-6 has been prepared by IEC technical committee 29: Electroacoustics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2009. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) the nominal test frequency used in DPOAE is now defined as the higher of the two frequencies, f_2 ;
- b) the permitted deviation of the stimulus signal for TEOAE has been specified;
- c) the frequency range for DPOAE stimulus signals has been redefined;
- d) the stimulus level requirements for TEOAE have been redefined;
- e) the stimulus level requirements for DPOAE have been redefined;
- f) the harmonic distortion requirements for DPOAE have been redefined;
- g) a minimum measurement range for DPOAE has been added.

The text of this International Standard is based on the following documents:

Draft	Report on voting
29/1109/FDIS	29/1114/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at <http://www.iec.ch/standardsdev/publications>.

A list of all parts in the IEC 60645 series, published under the general title *Electroacoustics – Audiometric equipment*, can be found on the IEC website.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

INTRODUCTION

Developments in the field of diagnostic hearing measurement have resulted in a number of instruments designed to evaluate the otoacoustic emissions of the human ear. Such emissions may be evoked by acoustic test signals having different spectral and temporal characteristics.

The practical use of such instruments concerns the measurement of sound energy emitted by the inner ear and its separation from sounds emerging from physiological or other sources.

The spontaneous otoacoustic emissions (SOAE) and stimulus frequency otoacoustic emissions (SFOAE), which comprise part of the otoacoustic emissions, are not covered by this document.

Conformance to the performance specification in this document is demonstrated when a measured deviation from a design goal equals or does not exceed the corresponding acceptance limit(s), and the laboratory has demonstrated that the associated uncertainty of measurement equals or does not exceed the maximum permitted uncertainty specified in this document.

INTRODUCTION to Amendment 1

The second edition of IEC 60645-6 was published in 2022 and included normative references to test signals of short duration specified in IEC 60645-3:2020. In 2025, IEC 60645-3:2020 was withdrawn, and its normative content was incorporated into IEC 60645-7:2025.

Consequently, IEC/TC 29 approved a project to develop Amendment 1 to IEC 60645-6:2022. This amendment includes

- replacing the references to the withdrawn IEC 60645-3:2020 with the relevant parts of IEC 60645-7:2025 concerning the requirements for short-duration signals, and
- editorial clarification of the two-sided ($\pm$) acceptance limits for the stimulus levels of DPOAE primary tones.

This amendment does not introduce any technical changes to the document.

1 Scope

This part of IEC 60645 applies to instruments designed primarily for the measurement of otoacoustic emissions in the human external auditory meatus evoked by acoustic probe stimuli. This document defines the characteristics to be specified by the manufacturer, specifies minimum mandatory functions for two types of instruments and provides performance specifications applicable to both instrument types. This document describes methods to be used to demonstrate conformance with the specifications in this document and guidance on methods for periodic calibration.

The purpose of this document is to ensure that measurements made under comparable test conditions with different instruments complying with this document will be consistent. Instruments can provide a measurement function not specifically within the scope of this document and still comply with the relevant requirements of this document for the functions that are within the scope. This document is not intended to restrict development or incorporation of new features, nor to discourage innovative approaches.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60318-4, *Electroacoustics – Simulators of human head and ear – Part 4: Occluded-ear simulator for the measurement of earphones coupled to the ear by means of ear inserts*

IEC 60318-5, *Electroacoustics – Simulators of human head and ear – Part 5: 2 cm³ coupler for the measurement of hearing aids and earphones coupled to the ear by means of ear inserts*

IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic disturbances – Requirements and tests*

IEC 60645-1:2017, *Electroacoustics – Audiometric equipment – Part 1: Equipment for pure-tone and speech audiometry*

~~IEC 60645-3:2020, *Electroacoustics – Audiometric equipment – Part 3: Test signals of short duration*~~

IEC 60645-7:2025, *Electroacoustics - Audiometric equipment - Part 7: Instruments for the measurement of auditory evoked potentials*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

Bibliography

IEC 60318-8, *Electroacoustics – Simulators of human head and ear – Part 8: Acoustic coupler for high-frequency measurements of hearing aids and earphones coupled to the ear by means of ear inserts*

ISO 389-6, *Acoustics – Reference zero for the calibration of audiometric equipment – Part 6: Reference threshold of hearing for test signals of short duration*
