

INTERNATIONAL STANDARD

Corrected version
2026-08

NORME INTERNATIONALE

**Ultrasonics – Hydrophones –
Part 1: Measurement and characterization of medical ultrasonic fields**

**Ultrasons – Hydrophones –
Partie 1: Mesurage et caractérisation des champs ultrasoniques médicaux**





IEC 62127-1

Edition 2.0 2022-03

INTERNATIONAL STANDARD

Corrected version
2026-08

NORME INTERNATIONALE

**Ultrasonics – Hydrophones –
Part 1: Measurement and characterization of medical ultrasonic fields**

**Ultrasons – Hydrophones –
Partie 1: Mesurage et caractérisation des champs ultrasoniques médicaux**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-5202-4

Warning! Make sure that you obtained this publication from an authorized distributor. Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.
--

INTERNATIONAL ELECTROTECHNICAL COMMISSION

IEC 62127-1
Edition 2.0 2022-03

Ultrasonics - Hydrophones -
Part 1: Measurement and characterization of medical ultrasonic fields

INTERPRETATION SHEET 1

This interpretation sheet has been prepared by IEC technical committee 87: Ultrasonics.

The text of this interpretation sheet is based on the following documents:

DISH	Report on voting
87/941/DISH	87/946/RVDISH

Full information on the voting for the approval of this interpretation sheet can be found in the report on voting indicated in the above table.

Subclause 5.1.7.1, last paragraph

This is clarified by the following:

Medical ultrasound devices often require additional acoustic field measurements during version upgrades to incorporate newly developed probes, modes, or transmission technologies into the system. In such cases, even within the same device, 5.1.7.1 permits data from a second edition conforming measurement system and data from earlier measurement systems to coexist depending on the testing period.

CONTENTS

FOREWORD	6
INTRODUCTION	8
1 Scope	9
2 Normative references	9
3 Terms and definitions	10
4 Symbols	32
5 Measurement requirements	34
5.1 Requirements for hydrophones and amplifiers	34
5.1.1 Preface	34
5.1.2 General	34
5.1.3 Sensitivity of a hydrophone	35
5.1.4 Directional response of a hydrophone	35
5.1.5 Effective hydrophone size	35
5.1.6 Choice of the size of a hydrophone active element	35
5.1.7 Bandwidth	37
5.1.8 Linearity	40
5.1.9 Hydrophone signal amplifier	40
5.1.10 Hydrophone cable length and amplifiers	40
5.2 Requirements for positioning and water baths	41
5.2.1 General	41
5.2.2 Positioning systems	41
5.2.3 Water bath	42
5.3 Requirements for data acquisition and analysis systems	43
5.4 Recommendations for ultrasonic equipment being characterized	43
6 Measurement procedure	43
6.1 General	43
6.2 Preparation and alignment	44
6.2.1 Preparation	44
6.2.2 Aligning an ultrasonic transducer and a hydrophone	44
6.3 Measurement	44
6.4 Analysis	44
6.4.1 Corrections for restricted bandwidth and spatial resolution	44
6.4.2 Uncertainties	44
7 Beam characterization	45
7.1 General	45
7.2 Primary pressure parameters	46
7.2.1 General	46
7.2.2 Peak-compressional acoustic pressure and peak-rarefactional acoustic pressure	47
7.2.3 Spatial-peak RMS acoustic pressure	47
7.2.4 Local distortion parameter	48
7.3 Intensity parameters derived from acoustic pressure	48
7.3.1 General	48
7.3.2 Intensity parameters using pulse-pressure-squared integral	49
8 Requirements for specific ultrasonic fields	52

8.1	General.....	52
8.2	Diagnostic fields	52
8.2.1	Simplified procedures and guidelines.....	52
8.2.2	Pulsed wave diagnostic equipment	52
8.2.3	Continuous wave diagnostic equipment	53
8.2.4	Diagnostic equipment with low acoustic output	54
8.3	Therapy fields	54
8.3.1	Physiotherapy equipment.....	54
8.3.2	High intensity therapeutic ultrasonic fields	55
8.3.3	Non-focused and weakly focused pressure pulses	55
8.4	Surgical fields	55
8.4.1	Lithotripters and pressure pulse sources for other therapeutic purposes	55
8.4.2	Low frequency surgical applications.....	56
8.5	Fields from other medical applications	56
9	Conformity statement.....	56
9.1	General.....	56
9.2	Maximum probable values.....	56
9.3	Sampling.....	57
Annex A (informative) General rationale.....		58
Annex B (informative) Hydrophones and positioning		60
B.1	General.....	60
B.2	Electrical loading considerations	60
B.3	Hydrophone signal amplifier.....	60
B.4	Hydrophone cable length and amplifiers.....	60
B.5	Transducer positioning.....	61
B.6	Alignment of hydrophones.....	62
B.7	Water bath lining material	62
B.8	Recommendations for ultrasonic equipment being characterized.....	62
B.9	Types of hydrophones.....	63
B.9.1	Ceramic needle hydrophones	63
B.9.2	PVDF needle hydrophones	63
B.9.3	PVDF membrane hydrophones	63
B.9.4	Fibre-optic and optic hydrophones	64
B.9.5	Relative performance of different types.....	65
B.10	Typical specification data for hydrophones.....	65
Annex C (informative) Acoustic pressure and intensity		66
Annex D (informative) Voltage to pressure conversion		68
D.1	General.....	68
D.2	Hydrophone deconvolution procedure	69
D.3	Converting the data between double-sided and single-sided spectra	70
D.4	Use of hydrophone calibration data	72
D.4.1	Calibration data interpolation	72
D.4.2	Calibration data extrapolation	72
D.4.3	Regularization filtering.....	73
D.5	Implication of the hydrophone deconvolution process on measurement duration	74
D.6	Validation of deconvolution implementation.....	75
Annex E (informative) Correction for spatial averaging.....		76

E.1	Linear and quasilinear fields	76
E.2	Linear fields, quasilinear fields, and broadband nonlinearly distorted waveforms	78
Annex F (informative) Acoustic output parameters for multi-mode medical ultrasonic fields in the absence of scan-frame synchronization		81
F.1	General.....	81
F.2	Current philosophy.....	81
F.3	Need for an alternative approach	82
F.4	Proposed approach.....	82
F.4.1	Alternative philosophy	82
F.4.2	Alternative parameters.....	83
F.5	Measurement methods.....	84
F.5.1	General	84
F.5.2	Peak pressures.....	84
F.5.3	Temporal-average intensity	84
F.5.4	Frequency	85
F.5.5	Power	85
F.6	Discussion	85
F.6.1	Relationship to existing standards	85
F.6.2	Advantages	86
F.6.3	Disadvantages.....	86
Annex G (informative) Propagation medium and degassing.....		87
Annex H (informative) Specific ultrasonic fields.....		88
H.1	Diagnostic fields	88
H.1.1	Useful relationships between acoustical parameters	88
H.1.2	Pulsed wave diagnostic equipment	89
H.1.3	Continuous wave diagnostic equipment	89
H.2	Therapy fields	90
H.2.1	Physiotherapy equipment.....	90
H.2.2	High intensity therapeutic ultrasonic equipment	90
H.2.3	Non-focused and weakly focused pressure pulses	90
H.3	Surgical fields	90
H.3.1	Lithotripters	90
H.3.2	Low frequency surgical applications.....	90
Annex I (informative) Assessment of uncertainty in the acoustic quantities obtained by hydrophone measurements.....		91
I.1	General.....	91
I.2	Overall (expanded) uncertainty	91
I.3	Common sources of uncertainty	91
Annex J (informative) Transducer and hydrophone positioning systems.....		93
Annex K (informative) Beamwidth midpoint method.....		94
Bibliography.....		95
Figure 1 – Schematic diagram of the different planes and lines in an ultrasonic field		12
Figure 2 – Several apertures and planes for a transducer of unknown geometry		26
Figure 3 – Parameters for describing a focusing transducer of known geometry.....		29
Figure 4 – Schematic diagram of the method of determining pulse duration		46
Figure D.1 – A flow diagram of the hydrophone deconvolution process		70

Figure D.2 – Example of waveform deconvolution	74
Figure J.1 – Schematic diagram of the ultrasonic transducer and hydrophone degrees of freedom	93
Table 1 – Acoustic parameters appropriate to various types of medical ultrasonic equipment.....	45
Table B.1 – Typical specification data for hydrophones, in this case given at 1 MHz [69]	65
Table C.1 – Properties of distilled or de-ionized water as a function of temperature [71]	67
Table D.1 – Method of conversion from a double- to a single-sided spectrum	71
Table D.2 – Method of conversion from a single- to a double-sided spectrum	71
Table F.1 – Main basic parameters defined in this document or in IEC 61161	82
Table F.2 – List of parameters that are to be used or are to be deleted.....	83
Table K.1 – Decibel beamwidth levels for determining midpoints.....	94

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – HYDROPHONES –

Part 1: Measurement and characterization of medical ultrasonic fields

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62127-1 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2007 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) The upper frequency limit of 40 MHz has been removed.
- b) Hydrophone sensitivity definitions have been changed to recognize sensitivities as complex-valued quantities.
- c) Procedures and requirements for narrow-band approximation and broadband measurements have been modified; details on waveform deconvolution have been added.
- d) Procedures for spatial averaging correction have been amended.
- e) Annex D, Annex E and bibliography have been updated to support the changes of the normative parts.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/783/FDIS	87/788/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

A list of all parts of IEC 62127 series, published under the general title *Ultrasonics – Hydrophones*, can be found on the IEC website.

NOTE Words in **bold** in the text are terms defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

The contents of the Interpretation sheet 1 (2026-08) have been included in this copy.

INTRODUCTION

The main purpose of this document is to define various acoustic parameters that can be used to specify and characterize ultrasonic fields propagating in liquids, and, in particular, water, using hydrophones. Measurement procedures are outlined that may be used to determine these parameters. Specific device related measurement standards, for example IEC 61689, IEC 61157, IEC 61847 or IEC 62359, can refer to this document for appropriate acoustic parameters. In IEC 62359, some additional measurement methods for attenuated parameters and indices are described addressing the specific needs of acoustic output characterization of ultrasonic diagnostic equipment in accordance with IEC 60601-2-37.

The philosophy behind this document is the specification of the acoustic field in terms of acoustic pressure parameters, acoustic pressure being the primary measurement quantity when hydrophones are used to characterize the field.

Intensity parameters are specified in this document, but these are regarded as derived quantities that are meaningful only under certain assumptions related to the ultrasonic field being measured.

ULTRASONICS – HYDROPHONES –

Part 1: Measurement and characterization of medical ultrasonic fields

1 Scope

This part of IEC 62127 specifies methods of use of calibrated **hydrophones** for the measurement in liquids of acoustic fields generated by ultrasonic medical equipment including **bandwidth** criteria and calibration frequency range requirements in dependence on the spectral content of the fields to be characterized.

This document:

- defines a group of acoustic parameters that can be measured on a physically sound basis;
- defines a second group of parameters that can be derived under certain assumptions from these measurements, and called derived intensity parameters;
- defines a measurement procedure that can be used for the determination of acoustic pressure parameters;
- defines the conditions under which the measurements of acoustic parameters can be made using calibrated **hydrophones**;
- defines procedures for correcting for limitations caused by the use of **hydrophones** with finite **bandwidth** and finite active element size, and for estimating the corresponding **uncertainties**.

NOTE 1 Throughout this document, SI units are used. In the specification of certain parameters, such as **beam areas** and intensities, it can be convenient to use decimal multiples or submultiples. For example, **beam area** is likely to be specified in cm^2 and intensities in W/cm^2 or mW/cm^2 .

NOTE 2 The **hydrophone** as defined can be of a piezoelectric or an optic type.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60565-1, *Underwater acoustics – Hydrophones – Calibration of hydrophones – Part 1: Procedures for free-field calibration of hydrophones*

IEC 61689, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 0,5 MHz to 5 MHz*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields up to 40 MHz*

IEC 62127-3, *Ultrasonics – Hydrophones – Part 3: Properties of hydrophones for ultrasonic fields up to 40 MHz*

IEC 63009, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 20 kHz to 500 kHz*

ISO 16269-6, *Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

Bibliography

IEC documents

IEC 60050-801, *International Electrotechnical Vocabulary – Part 801: Acoustics and electroacoustics*

IEC 60050-802, *International Electrotechnical Vocabulary – Part 802: Ultrasonics*

IEC 60500, *Underwater acoustics – Hydrophones – Properties of hydrophones in the frequency range 1 Hz to 500 kHz*

IEC 60601-2-5 *Medical electrical equipment – Part 2-5: Particular requirements for the basic safety and essential performance of ultrasonic physiotherapy equipment*

IEC 60601-2-37, *Medical electrical equipment – Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment*

IEC 60601-2-62, *Medical electrical equipment – Part 2-62: Particular requirements for the basic safety and essential performance of high intensity therapeutic ultrasound (HITU) equipment*

IEC TR 60854, *Methods of measuring the performance of ultrasonic pulse-echo diagnostic equipment*

IEC 61157, *Standard means for the reporting of the acoustic output of medical diagnostic ultrasonic equipment*

IEC 61161, *Ultrasonics – Power measurement – Radiation force balances and performance requirements*

IEC 61828, *Ultrasonics – Transducers – Definitions and measurement methods regarding focusing for the transmitted fields*

IEC 61846, *Ultrasonics – Pressure pulse lithotripters – Characteristics of fields*

IEC 61847, *Ultrasonics – Surgical systems – Measurement and declaration of the basic output characteristics*

IEC TS 61949, *Ultrasonics – Field characterization – In situ exposure estimation in finite-amplitude ultrasonic beams*

IEC 62359, *Ultrasonics – Field characterization – Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields*

IEC TS 62556, *Ultrasonics – Field characterization – Specification and measurement of field parameters for high intensity therapeutic ultrasound (HITU) transducers and systems*

IEC TR 62781, *Ultrasonics – Conditioning of water for ultrasonic measurements*

IEC 63045, *Ultrasonics – Non-focusing short pressure pulse sources including ballistic pressure pulse sources – Characteristics of fields*

IEC TS 63081, *Ultrasonics – Methods for the characterization of the ultrasonic properties of materials*

Other references

- [1] RADULESCU, EG., LEWIN, PA., NOWICKI, A. and BERGER, WA. Hydrophones' effective diameter measurements as a quasi-continuous function of frequency. *Ultrasonics*, 2003, vol. 41 iss. 8, p. 635-641
- [2] BEISSNER, K. Maximum hydrophone size in ultrasonic field measurements. *Acustica*, 1985, vol. 59, p. 61-66
- [3] HARRIS, GR. Transient field of a baffled planar piston having an arbitrary vibration amplitude distribution. *J. Acoust. Soc. Am.*, 1981, vol. 70, p. 186-204
- [4] FAY, B. Numerische Berechnung der Beugungsverluste im Schallfeld von Ultraschallwandlern (Numerical calculation of the diffraction losses in the sound field of ultrasonic transducers). *Acustica*, 1976, vol. 36, p. 209-213
- [5] KHIMUNIN, AS. On the ultrasound diffraction losses for circular transducers of different radii. *Acustica*, 1983, vol. 54, p. 13-22
- [6] BOUTKEDJIERT, T. and REIBOLD, R., Improvement of the lateral resolution of finite-sized hydrophones by deconvolution, *Ultrasonics*, 2000, vol. 38, p. 745-748
- [7] BOUTKEDJIERT, T. and REIBOLD, R., Reconstruction of ultrasonic fields by deconvolving the hydrophone aperture effects I. Theory and simulation, *Ultrasonics*, 2002, vol. 39, p. 631-639
- [8] BOUTKEDJIERT, T. and REIBOLD, R., Reconstruction of ultrasonic fields by deconvolving the hydrophone aperture effects II. Experiment, *Ultrasonics*, 2002, vol. 39, p. 641-648
- [9] HARRIS, GR. A model of the effects of hydrophone and amplifier frequency response on ultrasound exposure measurements. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, September 1991, vol. 38, iss 5, p. 413-417
- [10] HARRIS, GR. Are current hydrophone low frequency response standards acceptable for measuring mechanical/cavitation indices? *Ultrasonics*, 1996, vol. 34, iss. 6, p. 649-654
- [11] WILKENS, V., SONNTAG, S. and WEBER, M. Secondary complex-valued hydrophone calibration up to 100 MHz using broadband pulse excitation and a reference membrane hydrophone. *Proceedings of the IEEE Ultrasonics Symposium*, 2019, Glasgow, Scotland, (2019), WeH5.2
- [12] HOWARD, S. Calibration of Reflectance-Based Fiber-Optic Hydrophones, *Proceedings of the IEEE Ultrasonics Symposium*, 2016, DOI: 10.1109/ULTSYM.2016.7728572
- [13] EICHSTAEDT, S., WILKENS, V., HALE, P., DIENSTFREY, A., HUGHES, B. and JARVIS, C. On challenges in the uncertainty evaluation for time-dependent measurements. *Metrologia*, 2016, vol. 53, S125-S135
- [14] EICHSTAEDT, S. and WILKENS, V. Evaluation of uncertainty for regularized deconvolution: a case study in hydrophone measurements. *J. Acoust. Soc. Am.*, 2017, vol. 141. iss. 6, p. 4155-4167
- [15] HALLER, J., Wilkens, V., JENDERKA, K.-V. and KOCH, C. Characterization of a fiber-optic displacement sensor for measurements in HIFU fields, *J. Acoust. Soc. Am.*, 2011, vol. 129, no. 6, 3676-3681

- [16] WEAR, KA., LIU, Y., GAMMELL, PM., MARUVADA, S. and HARRIS, GR. Correction for frequency-dependent hydrophone response to nonlinear pressure waves using complex deconvolution and rarefactional filtering: application with fiber-optic hydrophones. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2015, vol. 62, iss. 1, p. 152-164
- [17] WILKENS, V., SONNTAG, S. and GEORG, O. Robust spot-poled membrane hydrophones for measurement of large amplitude pressure waveforms generated by high intensity therapeutic ultrasonic transducers. *J. Acoust. Soc. Am.*, 2016, vol. 139, iss. 3, p. 1319-1332
- [18] HURRELL, A, and RAJAGOPAL, S. The practicalities of obtaining and using hydrophone calibration data to derive pressure waveforms. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2017, vol. 64, iss. 1, p. 126-140
- [19] WEBER, M. and WILKENS, V. HIFU waveform measurement at clinical amplitude levels: primary hydrophone calibration, waveform deconvolution and uncertainty estimation. *Proceedings of the IEEE Ultrasonics Symposium*, 2017, DOI: 10.1109/ULTSYM.2017.8092385
- [20] WEAR, KA., GAMMELL, PM., MARUVADA, S., LIU, Y., and HARRIS, GR. Improved measurement of acoustic output using complex deconvolution of hydrophone sensitivity. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2014, vol. 61, no. 1, p. 62-74
- [21] EICHSTAEDT, S. and WILKENS, V. GUM2DFT – A software tool for uncertainty evaluation of transient signals in the frequency domain. *Meas. Sci. Technol.*, 2016, vol. 27, no. 5, 055001-1-055001-12
- [22] ISO/IEC Guide 98-3:2008/Suppl.2:2011, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995) – Supplement 2: Extension to any number of output quantities*
- [23] LEWIN, PA., SCHAFFER, ME. and CHIVERS, RC. Factors affecting the choice of preamplification for ultrasonic hydrophone probes. *Ultrasound Med. Biol.*, March 1987, vol. 13, no. 3, p. 141-148
- [24] PRESTON, RC., BACON, DR., LIVETT, AJ. and RAJENDRAN, K. PVDF membrane hydrophone performance properties and their relevance to the measurement of the acoustic output of medical ultrasonic equipment. *J. Phys. E: Sci. Instrum.*, 1983, vol. 16, p. 786-796
- [25] SMITH, RA. The importance of the frequency response of a hydrophone when characterising medical ultrasonic fields. *Proc. Inst. Acoustics*, 1986, vol. 8, Part 2, p. 119-128
- [26] LEWIN, PA. Calibration of Hydrophones. *Technical Review*, Brüel and Kjær, 1973, vol. 1, p. 3-17
- [27] PRESTON, RC., BACON, DR. and SMITH, RA. Calibration of medical ultrasonic equipment – procedures and accuracy assessment. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, March 1988, vol. 35 iss. 2, p. 110-121
- [28] HARRIS, GR. Hydrophone measurements in diagnostic ultrasound fields. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, March 1988, vol. 35, iss. 2, p. 87-101
- [29] SHOMBERT, DG. and HARRIS, GR. Use of miniature hydrophones to determine peak intensities typical of medical ultrasound devices. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, May 1986, vol. 33, iss. 3, p. 287-294

- [30] AIUM/NEMA. *Acoustic output measurement standard for diagnostic ultrasound equipment*. (NEMA Standards Publication UD 2-2004, Revision 3). Laurel, MD: American Institute of Ultrasound in Medicine (AIUM); Rosslyn, VA: National Electrical Manufacturers Association (NEMA), 2004
- [31] ZEQRIRI, B. and BOND, AD. The influence of waveform distortion on hydrophone spatial-averaging corrections – Theory and measurement. *J. Acoust. Soc. Am.*, 1992, vol. 92, p. 1809-1821
- [32] EC COMMUNITY BUREAU OF REFERENCE (BCR). *Development of standard measurement methods for essential properties of ultrasound therapy equipment* (TNO Report CMT/92.031). Leiden, The Netherlands: Centre for Medical Technology, 1992
- [33] HEKKENBERG, RT., REIBOLD, R. and ZEQRIRI, B. Development of standard measurement methods for essential properties of ultrasound therapy equipment. *UltrasoundMed.Biol.*, 1994, vol. 20, no. 1, p. 83-98
- [34] WILKENS, V. and MOLKENSTRUCK, W. Broadband PVDF membrane hydrophone for comparisons of calibration methods up to 140 MHz. *IEEE Trans. Ultrason., Ferroelect., Freq. Contr.*, 2007, vol. 54, p. 1784-1791
- [35] WEBER, M. and WILKENS, V. Using a heterodyne vibrometer in combination with pulse excitation for primary calibration of hydrophones in amplitude and phase. *Metrologia*, 2017, vol. 54, p. 432-444
- [36] WILKENS, V. and KOCH, C. Amplitude and phase calibration of hydrophones up to 70 MHz using broadband pulse excitation and an optical reference hydrophone. *J. Acoust. Soc. Am.*, 2004, vol. 115, p. 2892-2903
- [37] RADULESCU, EG., WÓJCIK, J., LEWIN, PA. and NOWICKI, A. Nonlinear propagation model for ultrasound hydrophones calibration in the frequency range up to 100 MHz. *Ultrasonics*, June 2003, vol. 41, iss. 4, p. 239-245
- [38] RADULESCU, EG., LEWIN, PA., WÓJCIK, J. and NOWICKI, A. Calibration of ultrasonic hydrophone probes up to 100 MHz using time gating frequency analysis and finite amplitude waves. *Ultrasonics*, June 2003, vol. 41, iss. 4, p. 247-254
- [39] STAUDENRAUS, J. and EISENMENGER, W. Fiber-optic hydrophone for ultrasonic and shockwave measurements in water. *Ultrasonics*, 1993, vol. 31, iss. 4, p. 267-273
- [40] LEWIN, PA., UMCHID, S., SUTIN, A. and SARVAZYAN, A. Beyond 40 MHz frontier: the future technologies for calibration and sensing of acoustic fields. *J. Phys.: Conf. Ser.*, 2004, vol. 1, p. 38-43
- [41] LEWIN, PA. and CHIVERS, RC. Two miniature ceramic ultrasonic probes. *J. Phys. E: Sci. Instrum.*, 1981, vol. 14, p. 1420-1424
- [42] SHOMBERT, DG., SMITH, SW. and HARRIS, GR. Angular response of miniature ultrasonic hydrophones. *Med. Physics*, July-August 1982, vol. 9, no. 4, p. 484-492
- [43] WEAR, KA., BAKER, C., and MILORO, P. Directivity and frequency-dependent effective sensitive element size of needle hydrophones: predictions from four theoretical forms compared with measurements. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2018, vol. 65, no. 10, 1781-1788
- [44] LEWIN, PA. Miniature piezoelectric polymer ultrasonic hydrophone probes. *Ultrasonics*, September 1981, vol. 19, iss. 5, p. 213-216

- [45] WILSON, DT., TANCRESS, RH. and CALLERAME, J. PVF2 polymer microprobe. *Proceedings of the IEEE Ultrasonics Symposium*, 1979, p. 506-510
- [46] BACON, DR. Characteristics of a PVDF membrane hydrophone for use in the range 1-100 MHz. *IEEE Transactions on Sonics and Ultrasonics*, January 1982, vol. SU-29, no. 1, p. 18-25
- [47] DEREGGI, AS., ROTH, SC., KENNEY, JM., EDELMAN, S. and HARRIS, GR. Piezoelectric polymer probe for ultrasonic applications. *J. Acoust. Soc. Am.*, 1981, vol. 69, p. 853-859
- [48] UEBERLE, F. Shockwave Measurements Using an Optical Light Spot Hydrophone. *BMT Conference Proceedings*, Zürich, September 2006
- [49] GRANZ, B., NANKE, R., FEHRE, J., PFISTER, T. and ENGELBRECHT, R. Light Spot Hydrophone, Innovation in Lithotripsy. *Medical Solutions*, June 2004, p. 86-87
- [50] JAMSHIDI RAD, A., UEBERLE, F., KRUGER, K. Investigation on the comparability of the light spot hydrophone and the fiber optic hydrophone in lithotripter field measurements. *Rev. Sci. Instr.*, 2014, vol. 85, 014902
- [51] KOCH, C. and REIBOLD, R. Interferometric fiber-optic sensor for measurement of lithotripter shock waves. *Proceedings of the 1995 World Congress on Ultrasonics*, 1995, p. 931-934. ISBN 3-9805013-0-2
- [52] KOCH, C. Measurement of ultrasonic pressure by heterodyne interferometry with a fiber-tip sensor. *Appl. Optics*, May 1999, vol. 38, iss. 13, p. 2812-2819
- [53] KOCH, C., MOLKENSTRUCK, W. and REIBOLD, R. Shock-wave measurement using a calibrated interferometric fiber-tip sensor. *Ultrasound Med. Biol.*, 1997, vol. 23, iss. 8, p. 1259-1266
- [54] BEARD, PC. and MILLS, TN. Extrinsic optical-fiber ultrasound sensor using a thin polymer film as a low-finesse Fabry-Perot interferometer. *Appl. Optics*, February 1996, vol. 35, iss. 4, p. 663-675
- [55] KOCH, C. Coated fiber-optic hydrophone for ultrasonic measurement. *Ultrasonics*, August 1996, vol. 34, iss. 6, p. 687-689
- [56] WILKENS, V.; KOCH, C. Fiber-optic multilayer hydrophone for ultrasonic measurement, *Ultrasonics*, 1999, vol. 37, p. 45-49
- [57] COLEMAN, AJ., DRAGUIOTI, E., TIPTAF, R., SHOTRI, N. and SAUNDERS, JE. Acoustic performance and clinical use of a fibreoptic hydrophone. *Ultrasound Med. Biol.*, January 1998, vol. 24, iss. 1, p. 143-151
- [58] UNO, Y. and NAKAMURA, K. Pressure sensitivity of a fibre-optic microprobe for high frequency ultrasonic field. *Jpn. J. Appl. Phys.*, 1999, vol. 38, p. 3120-3123
- [59] BEARD, PC., PERENNES, F. and MILLS, TN. Transduction mechanisms of the Fabry-Perot polymer film sensing concept for wideband ultrasound detection. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, November 1999, vol. 46, iss. 6, p. 1575-1582
- [60] BEARD, PC., HURRELL, AM. and MILLS, TN. Characterization of a polymer film optical fiber hydrophone for use in the range 1 to 20 MHz: A comparison with PVDF needle and membrane hydrophones. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, January 2000, vol. 47, iss. 1, p. 256-264

- [61] WURSTER, C., KÖHLER, M., PECHA, R., EISENMENGER, W., SUHR, D., IRMER, U., BRUMMER, F. and HÜLSER, D. Negative pressure measurements of water using the glass fiber optic hydrophone. *Proceedings of the 1995 World Congress on Ultrasonics*, 1995, p. 635-638. ISBN 3-9805013-0-2
- [62] WANG, ZQ., LAUXMANN, P., WURSTER, C., KÖHLER, M., GOMPF, B. and EISENMENGER, W. Impulse response of a fiber optic probe hydrophone determined with shock waves in water. *J. Appl. Phys.*, 1 March 1999, vol. 85, iss. 5, p. 2514-2516
- [63] KRÜCKER, JF., EISENBERG, A., KRIX, M., LÖTSCH, R., PESSEL, M. and TRIER, H. Rigid piston approximation for computing the transfer function and angular response of a fiber-optic hydrophone. *J. Acoust. Soc. Am.*, 2000, vol. 107, p. 1994-2003
- [64] WEISE, W., WILKENS, V. and KOCH, C. Frequency response of fiber-optic multilayer hydrophones: experimental investigation and finite element simulation. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, July 2002, vol. 49, iss. 7, p. 937-945
- [65] WILKENS, V. and KOCH, C. Optical multilayer detection array for fast ultrasonic field mapping. *Opt. Lett.*, 1999, vol. 24, 1026-28
- [66] BEARD, PC. Two-dimensional ultrasound receive array using an angle-tuned Fabry-Perot Polymer film sensor for transducer field characterization and transmission ultrasound imaging. *IEEE Trans. Ultrason. Ferroelec. Freq. Contr.*, 2005, vol. 52, pp. 1002-12
- [67] WILKENS, V. Characterization of an optical multilayer hydrophone with constant frequency response in the range from 1 to 75 MHz. *J. Acoust. Soc. Am.*, 2003, vol. 113, 1431-38
- [68] BEISSNER, K. The influence of membrane reflections on ultrasonic power measurements. *Acustica*, 1982, vol. 50, p. 194-200
- [69] LEWIN, PA. and SCHAFER, ME. Ultrasonic probes in measurement practice. *Medical Device and Diagnostic Industry*, 1986, vol. 8, no. 5, p. 40-45
- [70] BEISSNER, K. On the plane-wave approximation of acoustic intensity. *J. Acoust. Soc. Am.*, 1982, vol. 71, p. 1406-1411
- [71] KAYE, GWC. and LABY, TH. *Tables of Physical and Chemical Constants and Some Mathematical Functions*. 14th edition. Longman Group Ltd., 1973
- [72] KOCH, C. Amplitude and phase calibration of hydrophones by heterodyne and time-gated time-delay spectrometry. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, March 2003, vol. 50, iss. 3, p. 344-348
- [73] WEAR, KA., GAMMELL, PM., MARUVADA, S., LIU, Y., and HARRIS, GR. Time-delay spectrometry measurement of magnitude and phase of hydrophone response. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2011, vol. 58, no. 11, p. 2325-2333
- [74] BLOOMFIELD, PE. and LEWIN, PA. Determination of ultrasound hydrophone phase from Fourier-Hilbert transformed 1-40 MHz time delay spectrometry amplitude, *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2014, vol. 61, no. 4, p. 662-672
- [75] QUATERI, TF. and OPPENHEIM, AV. Iterative techniques for minimum phase signal reconstruction from phase of magnitude, *IEEE Trans. on acoustics, speech, and signal processing*, 1981, vol. ASSP-29, no. 6, p. 1187-1193

- [76] BECHHOFFER, J. Kramers-Kronig, Bode, and the meaning of zero, *Am. J. Phys.*, 2011, vol. 79, no. 10, p. 1053-1059
- [77] WEBER, M., WILKENS, V. and EICHSTAEDT, S. Tutorial for the deconvolution of hydrophone measurement data. DOI 10.5281/zenodo.4012242, <https://github.com/Ma-Weber/Tutorial-Deconvolution/tree/v1.0>
- [78] ORAN BRIGHAM, E. *The Fast Fourier Transform and its Applications*. Prentice Hall, 1988. ISBN 0 13-307547-8
- [79] WEAR, KA., LIU, Y. and HARRIS, GR. Pressure pulse distortion by needle and fiber-optic hydrophones due to nonuniform sensitivity. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2018, vol. 65, no. 2, p. 137-148
- [80] PRESTON, RC., BACON, DR. and SMITH, RA. Calibration of medical ultrasonic equipment-procedures and accuracy assessment. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, March 1988, vol. 35, iss. 2, p. 110-121
- [81] SMITH, RA. Are hydrophones of diameter 0.5 mm small enough to characterise diagnostic ultrasound equipment? *Phys. Med. Biol.*, 1989, vol. 34, p. 1593-1607
- [82] SELFRIDGE, A. and LEWIN, PA. Wideband Spherically Focused PVDF Acoustic Sources for Calibration of Ultrasound Hydrophone Probes. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, November 2000, vol. 47, iss. 6, p. 1372-1376
- [83] RADULESCU, EG., LEWIN, PA., GOLDSTEIN, A. and NOWICKI, A. Hydrophone spatial averaging corrections from 1-40 MHz. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, November 2001, vol. 48, iss. 6, p. 1575-1580
- [84] RADULESCU, EG., LEWIN, PA. and NOWICKI, A. 1-60 MHz Measurements in Focused Acoustic Fields using Spatial Averaging Corrections. *Ultrasonics*, May 2002, vol. 40, iss. 1-8, p. 497-501
- [85] LEWIN, PA., BARRIE-SMITH, N., HYNENEN, K., IDE, M. and MACDONALD, M. Interlaboratory Acoustic Power Measurement. *J. Ultrasound in Medicine*, 2003, vol. 22, p. 207-213
- [86] SMITH, RA. *The evaluation of small area hydrophones* [NPL Report RSA(EXT)2]. Teddington, Middlesex, UK: National Physical Laboratory, 1989
- [87] WEAR, KA. Considerations for choosing sensitive element size for needle and fiber-optic hydrophones – part I: spatiotemporal transfer function and graphical guide. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2019, vol. 66, no. 2, p. 318-339
- [88] WEAR, KA. and LIU, Y. Considerations for choosing sensitive element size for needle and fiber-optic hydrophones – part II: experimental validation of spatial average model. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2019, vol. 66, no. 2, p. 340-347
- [89] WEAR, KA., SHAH A., and BAKER C. Correction for hydrophone spatial averaging artifacts for circular sources. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2020, vol. 67, no. 12, p. 2674-2691
- [90] WEAR, KA., BAKER, C. and MILORO, P. Directivity and frequency-dependent effective sensitive element size of membrane hydrophones: theory vs. experiment. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2019, vol. 66, no. 11, p.1723-1730

- [91] WEAR, KA. and HOWARD, SM. Correction for spatial averaging artifacts in hydrophone measurements of high-intensity therapeutic ultrasound: an inverse filter approach. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, 2019, vol. 66, no. 9, p. 1453-1464
- [92] COOLING, M. P., HUMPHREY, V. F., AND WILKENS, V. Hydrophone area-averaging correction factors in nonlinearly generated ultrasonic beams. *J. Phys. Conf. Ser.* 279, Adv. Metrol. For Ultrasound in Med., 2011, 012002
- [93] HEKKENBERG, RT., BEISSNER, K., ZEQUIRI, B., BEZEMER, RA. and HODNETT, M. Validated ultrasonic power measurements up to 20 W. *Ultrasound Med. Biol.*, 2001, vol. 27, no. 3, p. 427-438
- [94] HEKKENBERG, RT., BEISSNER, K. and ZEQUIRI, B. *Therapy-level ultrasonic power measurement*. Final Technical Report, EC project SMT4-CT96-2139, European Commission report EUR 19510, 2000. ISBN 92-828-9027-9
- [95] HEKKENBERG, RT., BEISSNER, K. and ZEQUIRI, B. *Guidance on propagation medium and degassing for ultrasonic power measurements in the range of physiotherapy-level ultrasonic power*. EC project SMT4-CT96-2139, European Commission report EUR 19511, 2000. ISBN 92-828-9838-5
- [96] RADULESCU, EG., WÓJCIK, J., LEWIN, PA. and NOWICKI, A. Probing acoustic fields of clinically relevant transducers: the effect of hydrophone probes' finite apertures and bandwidths. *IEEE Trans. Ultrason. Ferroelectr. Freq. Contr.*, October 2004, vol. 51, iss. 10, p. 1262-1270
- [97] DUCK, FA. and STARRITT, HC. Acoustic shock generation by ultrasonic imaging equipment. *Brit. J. Radiol.*, 1984, vol. 57, p. 231-240
