

INTERNATIONAL STANDARD

REDLINE VERSION

HORIZONTAL PUBLICATION

**Safety requirements for electrical equipment for measurement, control, and
laboratory use -
Part 2-020: Particular requirements for laboratory centrifuges**

CONTENTS

FOREWORD	2
1 Scope and object.....	4
2 Normative references	5
3 Terms and definitions	5
4 Tests	7
5 Marking and documentation	7
6 Protection against electric shock	10
7 Protection against mechanical HAZARDS	10
8 Mechanical resistance to shock and impact Resistance to mechanical stresses.....	17
9 Protection against the spread of fire	17
10 Equipment temperature limits and resistance to heat	17
11 Protection against HAZARDS from fluids and solid foreign objects	17
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	19
13 Protection against liberated gases, explosion and implosion and escape of microbiological materials	19
14 Components and subassemblies	19
15 Protection by interlocks	19
16 HAZARDS resulting from application	19
17 RISK assessment	19
Annexes	20
Annex L (informative) Index of defined terms	20
Annex AA (normative) Dynamic microbiological test method for bioseals	22
Annex BB (informative) General guidance and rationale for particular subclauses	25
Annex CC (informative) General guidance for an empirical method to determine the kinetic energy of a ROTOR	30
Bibliography.....	32
Figure 101 CC.1 – ROTOR test setup	30
Table 101 – Time-temperature conditions	19

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**Safety requirements for electrical equipment for
measurement, control, and laboratory use -
Part 2-020: Particular requirements for laboratory centrifuges**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition IEC 61010-2-020:2016. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

IEC 61010-2-020 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment. It is an International Standard.

This fourth edition cancels and replaces the third edition published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with changes introduced by Amendment 1:2016 of IEC 61010-1:2010.

It has the status of a product safety publication in accordance with IEC Guide 104.

The text of this International Standard is based on the following documents:

Draft	Report on voting
66/820/CDV	66/868/RVC

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

This Part 2-020 is intended to be used in conjunction with the latest edition of IEC 61010-1. It was established on the basis of the third edition (2010) and its Amendment 1 (2016), hereinafter referred to as Part 1.

This Part 2-020 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for laboratory centrifuges*.

Where a particular subclause of IEC 61010-1 is not mentioned in this Part 2-020, that subclause applies as far as is reasonable. Where this Part 2-020 states "addition", "modification" or "replacement", the relevant requirement, test specification or note in IEC 61010-1 shall be adapted accordingly.

In this standard:

- the following print types are used:
 - requirements: in roman type;
 - NOTES: in small roman type;
 - *conformity and tests: in italic type*;
 - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS.
- subclauses, tables or figures which are additional to those in IEC 61010-1 are numbered starting from 101. Additional annexes are lettered starting from AA.

A list of all parts of the IEC 61010 series, published under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

1 Scope and object

IEC 61010-1:2010, Clause 1 and IEC 61010-1:2010/AMD1:2016, Clause 1 apply except as follows:

1.1.4 Scope

1.1.1 Equipment included in scope

Replacement:

This part of IEC 61010 is applicable to electrically powered LABORATORY CENTRIFUGES.

~~This group safety publication is primarily intended to be used as a product safety standard for the products mentioned in the scope, but shall also be used by technical committees in the preparation of its publications for products similar to those mentioned in the scope of this standard, in accordance with the principles laid down in IEC Guide 104 and ISO/IEC Guide 51.~~

~~NOTE—If~~ It is possible that all or part of the equipment falls within the scope of one or more other Part 2 standards of IEC 61010 as well as within the scope of this document, ~~it will also need to meet the requirements of those other Part 2 standards.~~ In that case, the requirements of those other Part 2 standards will also apply.

1.1.2 Equipment excluded from scope

Addition:

Add the following new item:

- aa) IEC 60034 series (Rotating electrical ~~machinery~~ machines).

1.2 Object

1.2.1 Aspects included in scope

Addition:

Add the following new items:

- aa) contact with moving parts (see 7.3);
- bb) LABORATORY CENTRIFUGE movement during any DISRUPTION (see 7.34.101);
- cc) high energy chemical reaction after ROTOR DISRUPTION (see 7.7.2.2 ~~k~~);
- dd) ineffectiveness of BIOSEALS (see 13.101).

1.2.2 Aspects excluded from scope

Addition:

Add the following new items:

- aa) additional precautions which ~~may need to be~~ are observed when centrifuging materials which are flammable or explosive (see 5.4.101);
- bb) additional precautions which ~~may need to be~~ are observed when centrifuging materials that could react chemically with sufficient vigour to cause a HAZARD (see 5.4.101).

1.4 Environmental conditions

1.4.1 Normal environmental conditions

Replacement:

Replace item c) with the following:

c) temperature 2 °C to 40 °C;

1.4.2 Extended environmental conditions

Replacement:

Replace item c) with the following:

c) ambient temperatures below 2 °C or above 40 °C;

2 Normative references

IEC 61010-1:2010, Clause 2 and IEC 61010-1:2010/AMD1:2016, Clause 2 apply except as follows:

Addition:

ISO 3864 (all parts), *Graphical symbols - Safety colours and safety signs*

Bibliography

Addition:

Add the following publications:

IEC 60034 (all parts), *Rotating electrical machines*

IEC 62061, *Safety of machinery - Functional safety of safety-related control systems*

ISO 13849 (all parts), *Safety of machinery - Safety-related parts of control systems*

ISO 19069-1:2015, *Plastics - Polypropylene (PP) moulding and extrusion materials - Part 1: Designation system and basis for specifications*

ISO 19069-2:2016, *Plastics - Polypropylene (PP) moulding and extrusion materials - Part 2: Preparation of test specimens and determination of properties*

EN 71-1:2011, *Safety of toys - Part 1: Mechanical and physical properties*

EN 12423:1999, *Polypropylene twines*

BS 4881:1993, *Specification for polypropylene film cords, lines and twines*

World Health Organization. *Laboratory Biosafety Manual, 2nd Edition. Geneva, 1993*

U.S. Department of Health and Human Services, Centers for Disease Control and Prevention and National Institutes OF Health. *Biosafety in Microbiological and Biomedical Laboratories, 4th Edition. Washington, 1999*

Health and Safety Executive, Advisory Committee on Dangerous Pathogens. *Categorisation of biological agents according to hazard and categories of containment, 4th Edition. London, 1995*

HARPER, G.J. Evaluation of sealed containers for use in centrifuges by a dynamic microbiological test method. *J. Clin. Pathol.* 1984, 37, pp. 1134-1139